



Ministry of Health and Quality of Life

MHPQ/M. DIS/08-09/Q23

June 2009

Dear Sir,

**ADDENDUM TO MHPQ/M. DIS/08-09/Q23**  
**(Supply of FBS Fluoride Oxalate Tubes, etc.)**  
**(Annual Requirements July 2009 – December 2010)**

Please refer to our Invitation to bid concerning the above.

2. The following clarifications regarding items 1 to 5 are brought to your attention:-

**Item 1: FBS tubes shall be non-vacuum, material of the tube can be either Polypropylene (PP), Polystyrene (PS) or Polyethylene Terephthalate (PET)**

**Item 2: Lithium Heparin Tube shall be non-vacuum, material of tube can be either Polypropylene (PP) Polysterene (PS) or Polyethylene Terephthalate (PET)**

**Item 3: EDTA Tube shall be vacuum and K2, material of tube to be plastic**

**Item 4: Plain tube shall be vacuum, material of tube can be either Polypropylene (PP), Polystyrene (PS) or Polyethylene Terephthalate (PET)**

**Item 5: Citrated Tube shall be non-vacuum, material of tube to be plastic.**

3. Any inconvenience caused is much regretted.

Yours faithfully,

**(H. Hauroo)**  
for Senior Chief Executive

-----  
Please acknowledge receipt of this addendum by return fax 210 7600.

**Name of company:.....**

**Date : .....**

**Signature : .....**