



Ministry of Health and Quality of Life

MHPQ/MDIS/08-09/Q17

07 September 2009

From: Senior Chief Executive, Ministry of Health and Quality of Life

To: The Director / Manager

.....  
.....  
.....

Dear Sir,

**ADDENDUM TO MHPQ/MDIS/08-09/Q17**  
**SUPPLY OF INTRAVENOUS CANNULA, ETC**  
**(REQUIREMENTS JULY 09- DEC 2010)**

Please refer to our invitation to bid concerning the above.

2. The following are brought to your attention:

- (i) Page 52 - Items 8 to 13 (disposable syringe etc.) **should read either luer lock or luer slip type instead of Luer Lock.**
- (ii) Page 53 - Item 20 should read **Disposable Syringe 20 ml with needle 21G x 1 1/2" either luer lock or luer slip type individually packed** instead of Disposable Syringe 20 ml with needle 20G x 1/2" individually packed luer lock.

3. Any inconvenience caused is regretted.

Yours faithfully,

**(H.Hauroo)**  
for Senior Chief Executive

.....  
Please acknowledge receipt of this addendum by return fax 210 7600

**Name of company:**.....

**Date** : .....

**Signature** : .....